

STATE-OF-THE-ART REVIEW

Management of Bleeding and Hemolysis During Percutaneous Microaxial Flow Pump Support



A Practical Approach

Charlotte J. Van Edom, MD,^a Mario Gramegna, MD,^b Luca Baldetti, MD,^b Alessandro Beneduce, MD,^c Thomas Castelein, MD,^d Dieter Dauwe, MD, PhD,^e Pascal Frederiks, MD,^a Gennaro Giustino, MD, PhD,^f Marc Jacquemin, MD, PhD,^g Stefan P. Janssens, MD, PhD,^a Vasileios F. Panoulas, MD, PhD,^{h,i} Janine Pöss, MD,^j Alexander Rosenberg, MD,^{h,i} Hannah A.I. Schaubroeck, MD,^k Benedikt Schrage, MD,^l Guido Tavazzi, MD, PhD,^m Thomas Vanassche, MD, PhD,^a Leen Vercaemst,ⁿ Dirk Vlasselaers, MD, PhD,^e Pascal Vranckx, MD, PhD,^o Jan Belohlavek, MD, PhD,^p Diana A. Gorog, MD, PhD,^{i,q} Kurt Huber, MD,^r Alexandre Mebazaa, MD, PhD,^s Bart Meyns, MD, PhD,ⁿ Federico Pappalardo, MD,^t Anna M. Scandroglio, MD, PhD,^b Gregg W. Stone, MD,^f Dirk Westermann, MD, PhD,^u Alaide Chieffo, MD, PhD,^c Susanna Price, MD, PhD,^{h,i,*} Christophe Vandenbriele, MD, PhD^{a,h,*}

ABSTRACT

Percutaneous ventricular assist devices (pVADs) are increasingly being used because of improved experience and availability. The Impella (Abiomed), a percutaneous microaxial, continuous-flow, short-term ventricular assist device, requires meticulous postimplantation management to avoid the 2 most frequent complications, namely, bleeding and hemolysis. A standardized approach to the prevention, detection, and treatment of these complications is mandatory to improve outcomes. The risk for hemolysis is mostly influenced by pump instability, resulting from patient- or device-related factors. Upfront echocardiographic assessment, frequent monitoring, and prompt intervention are essential. The precarious hemostatic balance during pVAD support results from the combination of a procoagulant state, due to critical illness and contact pathway activation, together with a variety of factors aggravating bleeding risk. Preventive strategies and appropriate management, adapted to the impact of the bleeding, are crucial. This review offers a guide to physicians to tackle these device-related complications in this critically ill pVAD-supported patient population. (J Am Coll Cardiol Intv 2023;16:1707-1720)

© 2023 The Authors. Published by Elsevier on behalf of the American College of Cardiology Foundation. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>).

From the ^aDepartment of Cardiovascular Diseases, University Hospitals Leuven, Leuven, Belgium; ^bCardiac and Cardiac Surgery Intensive Care Unit, IRCCS San Raffaele Scientific Institute, Milan, Italy; ^cInterventional Cardiology Unit, San Raffaele Scientific Institute, Milan, Italy; ^dCardiovascular Center, Onze-Lieve-Vrouweziekenhuis, Aalst, Belgium; ^eIntensive Care Medicine, University Hospitals Leuven, Leuven, Belgium; ^fDepartment of Cardiology, The Zena & Michael A. Wiener Cardiovascular Institute, Mount Sinai, New York, New York; ^gDepartment of Laboratory Medicine, University Hospitals of Leuven, Leuven, Belgium; ^hDepartments of Cardiology and Critical Care, Royal Brompton & Harefield Hospitals, Guy's and St. Thomas' NHS Foundation Trust, London, United Kingdom; ⁱFaculty of Medicine, National Heart and Lung Institute, Imperial College London, London, United Kingdom; ^jDepartment of Internal Medicine/Cardiology, Heart Center Leipzig at the University of Leipzig, Leipzig, Germany; ^kIntensive Care Unit, Ghent University Hospital, Ghent University, Ghent, Belgium; ^lDepartment of Cardiology, University Heart & Vascular Center Hamburg, Hamburg, Germany; ^mAnaesthesia and Intensive Care, Fondazione Policlinico San Matteo IRCCS, Pavia, Italy; ⁿDepartment of Cardiac Surgery, University Hospitals Leuven, Leuven, Belgium; ^oDepartment of Cardiology and Intensive Care Medicine, Jessa Ziekenhuis, Hasselt, Belgium; ^pSecond Department of Medicine-Department of Cardiovascular Medicine, First Faculty of Medicine, Charles University and General University Hospital, Prague, Czech Republic; ^qPostgraduate Medical School, University of Hertfordshire, Hertfordshire, United Kingdom; ^rDepartments of Cardiology and Intensive Care Medicine, Clinic Ottakring and Sigmund Freud University, Medical School, Vienna, Austria; ^sDepartment of Anesthesiology and Critical Care and

ABBREVIATIONS AND ACRONYMS

Hb	= hemoglobin
ICU	= intensive care unit
LDH	= lactate dehydrogenase
LV	= left ventricle/ventricular
MCS	= mechanical circulatory support
NO	= nitric oxide
PAPI	= pulmonary artery pulsatility index
PAWP	= pulmonary artery wedge pressure
pfHb	= plasma-free hemoglobin
pVAD	= percutaneous ventricular assist device
RV	= right ventricle/ventricular
UFH	= unfractionated heparin

The use of percutaneous ventricular assist devices (pVADs) in the management of cardiogenic shock or as a bridge providing hemodynamic support during complex coronary procedures has significantly increased over the past decade, likely because of improved device design and better understanding of device management.¹ Among mechanical circulatory support (MCS) devices, microaxial pVADs such as the Impella (Abiomed) are commonly used.^{1,2} The Impella supports the left ventricle (LV) and/or the right ventricle (RV) by transferring blood across the aortic or tricuspid and pulmonary valves, on the basis of the principle of Archimedes' screw. It augments systemic and/or pulmonary forward flow, maintaining end-organ perfusion and also unloads the ventricle, which results in a reduced area inside the ventricular pressure-volume loop and consequently reduced myocardial oxygen demand.³

Whereas the duration of indwelling pVAD support was previously on the order of days, patients may now be supported for weeks because of extended unloading, axillary implantation with de-escalation strategies, and weaning of pVAD as the last step in recovery.³ However, despite superior hemodynamic support compared with the intra-aortic balloon pump, 3 large retrospective registries in the United States, including >5,000 patients supported by microaxial pVADs, did not demonstrate a survival benefit of using pVAD compared with intra-aortic balloon pumps.^{1,4,5} This was attributed mainly to a higher rate of major bleeding complications in the pVAD group, although none of these studies mentioned a standardized intensive care unit (ICU) management strategy.⁶ Furthermore, the retrospective design renders propensity-matched studies subject to selection bias that can affect outcomes. Another explanation for the disappointing outcomes with microaxial pVAD support might be the frequent occurrence of hemolysis, with a reported cumulative rate up to 62.5%.^{7,8}

HIGHLIGHTS

- Bleeding and hemolysis impair outcomes during pVAD support for cardiogenic shock.
- Prevention, adequate diagnostics, and management may mitigate these complications.
- Prospective studies are needed to standardize hematological management during MCS.

As bleeding and hemolysis remain important complications, the accurate prevention, monitoring, and management of these complications is pivotal when managing a pVAD-supported patient in the ICU (**Central Illustration**).⁹ Here, we discuss the mechanisms that underlie hemolysis and bleeding and vascular complications during percutaneous microaxial flow pump support and their prevention and optimal management.

MECHANISMS OF HEMOLYSIS AND BLEEDING

HEMOLYSIS. Hemolysis, the release of hemoglobin into the plasma from erythrocytes, leads to a decline in efficient oxygen delivery and may be detected as an increase in plasma-free hemoglobin (pfHb), which occurs when the capacity of protective hemoglobin-scavenging mechanisms (eg, haptoglobin) becomes saturated.¹⁰ PfHb consumes nitric oxide (NO), resulting in vasoconstriction, platelet activation and aggregation, and arterial thrombosis.¹⁰ Furthermore, pfHb also increases inflammation and can induce pigment nephropathy by precipitation, leading to acute kidney injury (**Figure 1**).^{11–13} Although the impact of hemolysis on mortality in patients on MCS is unclear, it is important to detect hemolysis in a timely fashion, in order to prevent acute kidney injury and other pfHb-induced complications.⁷

Hemolysis is a potential complication of all MCS devices and can be pump or patient related or a


Burn Unit, Saint-Louis and Lariboisière Hospitals, Paris, France; ¹Cardiothoracic and Vascular Anesthesia and Intensive Care, Azienda Ospedaliera Nazionale Santi Antonio e Biagio e Cesare Arrigo, Alessandria, Italy; and the ²Department of Cardiology and Angiology, University Heart Center Freiburg-Bad Krozingen, Faculty of Medicine, University of Freiburg, Freiburg, Germany. * Drs Price and Vandenbriele contributed equally to this work.

The authors attest they are in compliance with human studies committees and animal welfare regulations of the authors' institutions and Food and Drug Administration guidelines, including patient consent where appropriate. For more information, visit the [Author Center](#).

Manuscript received March 9, 2023; revised manuscript received May 9, 2023, accepted May 23, 2023.

CENTRAL ILLUSTRATION Hemolysis and Bleeding During Microaxial pVAD Support

Management and Time Course of Hemolysis and Bleeding During Microaxial pVAD Support




Van Edom CJ, et al. J Am Coll Cardiol Interv. 2023;16(14):1707-1720.

(A) As bleeding and hemolysis remain important complications during microaxial percutaneous ventricular assist device (pVAD) support, accurate management using a stepwise, standardized approach is crucial. **Green color** indicates prevention. **Yellow color** indicates measures for minor bleeds. **Orange** and **red colors** indicate additional measures for intermediate and major bleeds, respectively. (B) Bleeding complications typically occur in the first 2 to 3 days after device implantation and rise in parallel with the duration of the microaxial pVAD support. Thrombotic complications typically rise with the duration of support. Hemolysis is mostly present shortly after implantation and can be avoided by proper device positioning and management. BBPS = bicarbonate-based purge solution; RV = right ventricular; TXA = tranexamic acid; UFH = unfractionated heparin.

combination of both. As discussed later, various patient- and/or ICU-related issues may also strongly affect its occurrence.

Device-related factors. First, inherent function-related mechanisms of the microaxial pVAD might

contribute to the occurrence of hemolysis. Increased shear stress during pVAD support is an important factor resulting in erythrocyte damage. When the erythrocyte membrane is mechanically stressed, the cell's capacity to deform and perform its normal

FIGURE 1 Systemic Repercussions of (Acute) Intravascular Hemolysis

functions starts to decline. This results in hemolysis or nonreversible subhemolytic damage.¹⁴ The computational fluid dynamics of an Impella CP show higher shear stress at the tip of the impeller blade (between the rotor and the housing), which is in line with the established formula whereby linear pump speed and, consequently, shear stress are the highest at the outer tip of the impeller blades¹⁵:

$$v \text{ (linear speed)} = r \text{ (radius)} \times \omega \text{ (rotational speed)}$$

To keep impeller tip speed, along with the shear stress, below a defined level, pumps with larger

diameters are designed to generate flow with fewer rotations per minute. For example, the Impella 2.5 allowed operation up to 51,000 rotations per minute, while the Impella 5.0 allows only up to 33,000 rotations per minute, albeit with doubled output. Therefore, it is easily understandable that using a lower rotational speed in a microaxial pVAD reduces hemolysis.¹⁶ In a case series of 23 left Impella-supported patients (7×5.0 , $16 \times \text{CP}$), pfHb levels were indeed higher at high pump support levels (P7 and P8) compared with pfHb of patients on lower levels (P5 and P6). Furthermore, the investigators showed a

TABLE 1 Parameters to Obtain Proper Positioning of the Impella Device

Routine preimplantation assessment of LV dimensions and aortic and mitral annulus
Ensure unobstructed microaxial pVAD inflow
Device positioned ~3.5 cm below the aortic valve (Impella CP and 5.0) or ~5 cm (Impella 5.5); Impella RP inlet in the inferior vena cava
Pump housing in the midventricular cavity
Free from anterior mitral leaflet
Free from subannular structures
Tip of the catheter pointing toward LV apex
Aim for pVAD outflow well above the aortic valve (Impella CP and 5.0/5.5) and 2–4 cm above the pulmonary valve annulus (Impella RP)
Aim for a stable device position (avoid pump migration); consider axillary insertion for prolonged use in a mobile patient

Proper positioning is key to reduce the risk of hemolysis and bleeding.
LV = left ventricular; pVAD = percutaneous ventricular assist device.

direct correlation between P level and pfHb (Pearson's $R = 0.61$, $P = 0.0004$).¹⁷ Another unique feature of the Impella is the purge solution, which prevents the entrance of blood into the motor, as this would inevitably lead to hemolysis and pump thrombosis. Unfractionated heparin (UFH) is conventionally used in the dextrose purge solution to protect against adsorption, deposition, and coagulation of blood components. When the purge flow is obstructed (by, eg, thrombus), hemolysis will occur immediately. In addition, reduced purge flow may be insufficient to attenuate motor heat resulting from attrition forces, resulting in increased local temperature and higher risk for clotting and hemolysis.¹⁸

Second, hemolysis may also be related to (transient) suboptimal intracardiac device positioning (Table 1), resulting in partial inlet or outlet obstruction and suction events. Indeed, 50% obstruction of an Impella CP outlet leads to a detrimental increase in the exposure time of blood to regions of high shear stress because of flow restriction and increase in turbulence near the impeller and the outlet windows.¹⁵

Patient-related factors. Patient anatomical parameters can contribute to the feasibility of proper positioning of the pump. As recently showed by Nakamura *et al*,¹⁹ a narrow angle of $<126.5^\circ$ between the aortic and mitral annulus in the apical 3-chamber view on transthoracic echocardiography was an independent risk factor for refractory hemolysis in a group of 26 patients. The group with the narrow angle ($n = 11$) also showed significantly smaller left ventricular (LV) end-diastolic dimensions. This narrow angle oriented the microaxial pVAD toward the lateral or posterior wall and resulted in mechanical suction and obstruction due to improper positioning of the device. Consequently, the anatomical variability among patients should also be considered upon microaxial pVAD insertion, and a quick

echocardiographic assessment of the heart before device selection is mandatory to avoid hemolytic complications afterward. Furthermore, hemodynamic issues might lead to the occurrence of hemolysis. For example, low LV preload due to hypovolemia, vasoplegia, or RV failure can result in suction events with inflow obstruction and hemolysis. This was illustrated in a retrospective study showing that patients with RV impairment, indicated by lower pulmonary artery pulsatility index (PAPi) (<1.3) following the initiation of microaxial pVAD support, have a higher risk for hemolysis.²⁰ Finally, other factors, including blood transfusions, primary hemolytic disorders such as sickle cell anemia or autoimmune hemolytic anemia, surgery, and tissue damage, among others, can also contribute to clinically significant hemolysis not directly related to the pVAD itself.

BLEEDING. The most common cause of morbidity and mortality during pVAD support are bleeding and thrombotic complications. This is the consequence of a complex, bidirectional interplay among different factors influencing the precarious hemostatic balance.²¹ Indeed, during critical illness, a procoagulant state is often seen because of multiorgan failure and a systemic inflammatory response.²¹ Also the nonbiological material of the device itself will activate not only the intrinsic (contact) pathway of coagulation but also platelets and leukocytes, resulting in increased thrombotic risk. In contrast, different factors can aggravate bleeding risk.

First, the device-induced high-shear environment and reduction in pulsatility both lead to increased cleavage of and reduced endothelial release of high-molecular weight von Willebrand factor, resulting in acquired von Willebrand syndrome and, consequently, reduced platelet-binding affinity.²² This phenomenon might be facilitated by improper

intracardiac device position and orientation (Table 1), as suggested by a recent study demonstrating a significantly increased rate of clinically relevant bleeding events in patients with Impella malrotation inside the LV cavity.²³ Malrotation of the microaxial pVAD is defined by catheter inflow orientation away from the LV apex and toward the mitral valve apparatus and the LV inferolateral wall (Figure 2).

Second, to avoid device-related thrombotic complications, systemic anticoagulation, usually with UFH, is necessary but increases bleeding risk. Its dose management is far from easy, involving on-off phenomenon in case of bleeding further aggravating coagulopathy, UFH monitoring tests, and other issues.²² Different tests to monitor anticoagulation and detect bleeding risk in this patient cohort have been reviewed elsewhere.²⁴ Next to anticoagulant therapy, patients on pVAD therapy often require antiplatelet therapy for concomitant conditions, further affecting bleeding risk.

Third, the unique purge system can further complicate anticoagulant management; the purge solution is frequently heparinized, and its flow rate is automatically controlled by the device to maintain purge pressure within a predefined range.²¹

Fourth, the development of liver failure with depletion of coagulant factors II, V, VII, IX, and X again aggravates bleeding risk. Also, the microaxial pVAD may be run in conjunction with other extracorporeal circuits (eg, continuous renal replacement therapy, venoarterial extracorporeal circulation and blood purification devices), increasing the blood-device interaction and favoring platelet consumption. Finally, larger bore femoral access is required, increasing the risk for bleeding (and vascular complications). In particular, the insertion sheath is slightly larger than the repositioning sheath, which might also contribute to the risk for access-site bleeding.²⁵

In conclusion, the complex interplay among procoagulant and antithrombogenic factors, including patient-, pump-, and drug-related variables, substantially increases the risk for bleeding, vascular, and thrombotic complications during microaxial pVAD use.


DIAGNOSIS OF HEMOLYSIS AND BLEEDING

HEMOLYSIS. Clinical as well as biochemical parameters and technical tools are useful in suspecting and confirming hemolytic events as early as possible.

Clinical manifestations include a change in urine color due to hemoglobinuria (“tea colored urine”) and late jaundice. Biochemical changes in case of hemolysis include increases in pFHb, lactate dehydrogenase (LDH), phosphate, unconjugated bilirubin, and potassium levels and decreases in hemoglobin and haptoglobin.²⁶ Importantly, preanalytical error should always be ruled out (eg, delayed handling in the laboratory). All these biochemical parameters have limitations, pertaining to the timeliness or availability of the results and specificity for detecting hemolysis. For example, pFHb may be increased because of surgery and transfusion, LDH is a nonspecific marker of tissue damage, and the production of haptoglobin can be reduced because of liver damage. Clearly, all of these scenarios may occur in patients with cardiogenic shock, independent of pVAD support. Because of this, whichever biomarker is used, the threshold for hemolysis that represents clinically or prognostically relevant hemolysis is unclear. Although the Extracorporeal Life Support Organization considers a pFHb concentration >50 mg/dL as a cutoff for serious hemolysis, recent pVAD registries have used a cutoff of >40 mg/dL.^{27,28} A recent consensus document providing definitions for adverse events in patients on pVAD defines hemolysis as a pFHb concentration >20 mg/dL.²⁹ Others use a decrease in hemoglobin, blood transfusion requirement plus a decrease in haptoglobin, or an increase in LDH.⁷ A retrospective registry showed that an increase in pFHb of >27 mg/dL within 24 hours of pVAD implantation was superior to plasma LDH level in predicting hemolysis, with sensitivity of 57% and specificity of 93%.¹⁷ Pragmatically, the pFHb levels should be as low as possible, and any rise should prompt immediate evaluation. Therefore, frequent biochemical sampling, at least daily but more often when hemolysis is already present, is encouraged. Once hemolysis is suspected, immediate action should be undertaken.

Next, as inadequate microaxial pVAD positioning will inevitably lead to hemolysis, the aid of technical tools to assess dislocation of the pump is an important pillar in hemolysis prevention and detection. First, echocardiography is important to visualize the pump. The distance between the aortic valve and the inflow cage should preferably be approximately 3.5 cm for an Impella CP or 5 cm for an Impella 5.0/5.5. The Doppler color artifact that locates the outflow opening of the pump can be used to determine (dis) location of the pump. In addition, the observation of

FIGURE 2 Device-Anatomy Interactions Relevant for Bleeding and Thrombosis Caused by Malrotation



Impella malrotation within the LV (**upper panels**) generates device-anatomy interactions that might cause bleeding or thrombosis. Multiplanar reconstructions of computed tomographic images (**A to E**) illustrate interactions of device inflow with the mitral valve (MV) apparatus (**arrow**) and MV leaflets (**arrowhead**), shaft impingement on the aortic valve (AV) cusps (**dashed arrow**), and pigtail/inflow impingement on the LV lateral or inferolateral wall (*). Ao = aorta; LA = left atrium; LV = left ventricle/ventricular; RA = right atrium; RV = right ventricle.


bubble artifacts in the LV during echocardiography is a further hint that relevant hemolysis is present.

Second, fluoroscopy and computed tomography can be helpful to confirm dislocation, for both left-sided and right-sided pumps. Last, the placement and the motor current signals on the pVAD console should be used to continuously assess pump positioning. The placement signal registers the difference between the inlet and outlet pressure (for the Impella 5.0) or the aortic pressure (for the Impella CP) during the cardiac cycle and should be accompanied by a pulsatile motor current. This specifically indicates that the inlet and the outlet are positioned in different compartments (aorta or pulmonary artery vs ventricle), which leads to variations in flow through the microaxial pVAD and therefore variations in energy consumption of the device during systole and diastole. The only exception that leads to complete lack of pulsatility of the placement signals, despite good positioning, is when there is ventriculoarterial

uncoupling with consequent continuous closure of the aortic valve, and all blood being diverted through the pump.

In 2019, the SmartAssist technology was added to the Impella device to allow better positioning and avoidance of suction and, thus, hemolysis. By adding an optical placement sensor at the outlet of the pump, aortic pressure can be detected. This, in combination with registration of variability in microaxial motor current, which is proportionally related to the differential pressure between the aorta and LV, allows calculation of the LV pressure waveform. The display of the LV pressure waveform is informative to promptly detect suction, before the development of manifest hemolysis, as it can help differentiate between inadequate volume status and improper microaxial pVAD positioning. In particular, if the LV waveform shows negative diastolic pressures with normal systolic pressures, the filling and the volume status should be checked. In contrast, if the LV shows

FIGURE 3 SmartAssist Technology Allows Differentiation Between Hypovolemia and Impaired Device Positioning



both negative diastolic and systolic pressures, a positioning problem is more likely (Figure 3).

BLEEDING. Meticulous monitoring of the anticoagulant effect of UFH during pVAD support is key to minimizing the risk for bleeding. This is challenging, as different coagulation tests are affected by various factors during critical illness and pump support. In a recent review, our group gave an overview of the (dis)advantages of different coagulation tests and proposed an anti-Xa/activated partial thromboplastin time-driven algorithm to assess the anticoagulant effect of UFH in this critically ill patient population.^{21,22} Bleeding jeopardizes patients with pVADs, as it results in both direct harm, as well as need for transfusions and eventually cessation of antithrombotic therapy, which may ultimately lead to an increased risk for thrombotic and ischemic events. Although there is general consensus on endpoint definitions for ischemic events, the classification of bleeding is far from uniform.^{30,31} This

has resulted in a wide variability in the reported incidence of bleeding and undermines the assessment of different bleeding mitigation strategies in clinical trials. To harmonize the definition of bleeding, different scores have been proposed in different settings. The most commonly used bleeding scores and the settings for which they were developed are listed in Table 2, along with their major limitations in the ICU population.^{32–36} Given the frequency of bleeding and thrombosis in this predisposed patient cohort, a specific bleeding score designed for critically ill patients supported by pVADs is urgently needed.²⁹

APPROACH AND MANAGEMENT OF HEMOLYSIS AND BLEEDING

HEMOLYSIS. Once hemolysis is suspected, immediate action should be taken. We therefore propose the following practical approach, which combines investigation of the etiology of hemolysis with

TABLE 2 The Most Commonly Used Bleeding Classification Scores With Their Original Development Cohort and Major Limitations in the ICU Setting

Name of Bleeding Score	Field of Indication for the Bleeding Score
Bleeding Academic Research Consortium	Developed for cardiovascular clinical trials of antithrombotic therapy, derived from a cohort of mainly patients with acute coronary syndrome
Valve Academic Research Consortium	Developed for clinical trials in TAVR
International Society on Thrombosis and Haemostasis	Developed for long-term oral anticoagulation exposure
Thrombolysis In Myocardial Infarction	Developed for fibrinolytic therapy after STEMI/non-CABG-related bleeding
Major Limitations in the ICU Setting	
None of the scores was designed to assess bleeding on short-term intravenous anticoagulation. No incorporation of patient's baseline Hb, which is often around the transfusion threshold in critically ill patients. Consequently, transfusions may be administered to correct pre-existing deficiencies rather than in reaction to bleeding, therefore potentially misclassifying the severity of a minor bleed. No incorporation of reperfusion of bleed on patient hemodynamics (eg, increased pressor requirements). No weighting in the scores between severity and chronicity of bleeding (eg, chronic oozing around line insertion site or ENT area vs acute pulmonary hemorrhage). Incorporation of endpoints determined by the response to the bleed (eg, surgical intervention) that suffer from reduced standardization and are not directly applicable to an ICU population ("surgical intervention" poorly classified; could be interpreted to include both major and minor surgical interventions (eg, resectionotomy or an additional suture around an arterial or venous access site). Classification of bleed sometimes based on "actionable" vs "nonactionable": the threshold for intervention in the ICU setting may be much higher or much lower than the threshold for intervention for the same bleed outside the ICU setting. However, this allows capture of events outside arbitrary laboratory definitions. Incorporation of Hb levels into scores, variable laboratory cutoffs, and consequent uncertainty regarding the timing of assessment may lead to inappropriate Hb peaks and nadirs. Furthermore, alternative reasons for drop in Hb (eg, hemolysis, often present on MCS) are not taken into account. Built on consensus rather than derived from an analysis of any specific database to identify independent risk factors.	
CABG = coronary artery bypass grafting; ENT = ear, nose, and throat; Hb = hemoglobin; ICU = intensive care unit; MCS = mechanical circulatory support; STEMI = ST-segment elevation myocardial infarction; TAVR = transcatheter aortic valve replacement.	


general measures to avoid further complications (Figure 4).

First, proper device positioning should be checked, as previously explained (bedside transthoracic echocardiography, SmartAssist technology). If the position of the pump is not established yet, fluoroscopy, chest radiography, or computed tomography can be diagnostic. In case the device is situated too deep in the ventricle, repositioning (pull back), guided by transthoracic echocardiography, is the first bedside option. Importantly, the P level should be lowered to P2 before the pump is manipulated to avoid structural cardiac damage. If the microaxial pVAD has been dislodged above the aortic valve, advancing it at the bedside is generally not recommended.

Second, when pump malpositioning has been excluded on transthoracic echocardiography, RV function and ventricular preload should be evaluated. The LV pressure waveform on the console can again be informative in this scenario, as normal systolic and low diastolic pressures indicate low filling status. Moreover, pulmonary artery catheterization allows the determination of right atrial pressure, pulmonary artery wedge pressure (PAWP), and calculation of PAPI. Such information can help

differentiate between hypovolemia (low PAWP) and low central venous pressure and RV failure (low to normal PAWP and disproportionately high right atrial pressure, with PAPI <1.0) and pulmonary hypertension.³ A dynamic positive end-expiratory pressure test (lowering or increasing the positive end-expiratory pressure to assess its effect on hemodynamic parameters) and/or fluid challenge might also offer valuable information. If RV dysfunction is diagnosed, inotropes or pulmonary vasodilators or the addition of RV MCS should be promptly considered. When hypovolemia is presumed to be the main cause of insufficient pump preload, an empirical but cautious fluid bolus should be administered. In case of hypovolemia, it will result in reduced suction events and increased cardiac output. This scenario is usually driven by low cardiac output as the main clinical issue, and management should focus on hemodynamic stabilization; thereafter, pump position should be re-evaluated according to the modified geometry and pressures in the heart chambers.

Third, if device dislocation, RV dysfunction, and hypovolemia are excluded as the drivers of hemolysis, pump thrombosis should be considered, as a clot


FIGURE 4 Practical Step-by-Step Approach for Hemolysis Management During Microaxial Percutaneous Ventricular Assist Device Support

will narrow the lumen between the impeller and the housing and thereby enhance shear stress and induce hemolysis. Evidence for pump thrombosis includes elevated purge pressures and a recent history of insufficient anticoagulation, including an anti-Xa level <0.2 IU/L.²² Pump exchange is recommended when pump thrombosis is suspected.²² Finally, if the

cause of hemolysis remains unclear, more unusual etiologies, including intrinsic reasons for hemolysis (eg, sickle cell anemia) or thrombotic microangiopathy, should be considered.

Together with the etiologic investigation, both pump and ICU management-related general actions should be considered to lower the shear stress and

FIGURE 5 Prevention and Management of Bleeding Complications on Microaxial pVAD Support




Overview of strategies to prevent bleeding complications and suggested treatment modalities in case of minor, intermediate, or major bleeding. Source control should be central, aided by additional interventions depending on the type, location, and severity of the bleed. Furthermore, general measures based on maintenance of hemostasis and monitoring of coagulation tests are provided. BBPS = bicarbonate-based purge solution; ENT = ear, nose, and throat; GI = gastrointestinal; ICB = intracranial bleed; INR = international normalized ratio; IV = intravenous; pVAD = percutaneous ventricular assist device.

avoid direct complications of hemolysis. A first step is lowering the P level as much as possible, as this will reduce shear stress. If lower P levels are well tolerated, early weaning from microaxial pVAD should be attempted. Alternatively, if lower support is not sufficient, pump exchange or upgrade to another MCS should be considered. Upgrading patients on Impella CP support experiencing refractory hemolysis to a surgically implanted Impella device because of better fixation of the pump position was suggested.³⁷ As general management, prevention of hypovolemia and increased anticoagulation levels may be necessary. Counteracting the devastating effects of pfHb during excessive hemolysis is equally important. In the presence of severe pfHb-induced vasoconstriction, irreversible pfHb-NO binding may be countered by administering high doses of inhaled NO.^{38,39} Ideally, pfHb should be removed from the systemic

circulation as soon as possible to counter thrombosis, renal failure, vasoconstriction, and other adverse effects. This can be done by plasmapheresis (against 50% albumin, 50% fresh-frozen plasma) and/or by hemoadsorption (eg, CytoSorb, CytoSorbents).⁴⁰

BLEEDING. The most important step if bleeding occurs during pVAD support is to obtain bleeding source control (Figure 5). As most bleeds are access site related, local measures are the first step. Meticulous (ultrasound-guided) cannulation techniques, proper positioning, and stitching of the sheath to the skin are key to avoid complications (pull-up technique; Figure 6). Manual pressure and/or local application of gauze soaked in tranexamic acid or adrenaline (concentration 1:1,000 with a limited application time of 20 minutes to avoid necrosis of the skin) is often effective. For ear, nose, and throat bleeds, local treatments are also recommended (eg, mouth packing

FIGURE 6 Measures to Ensure Proper Positioning of the Femoral Impella Sheath**Proper positioning of the femoral Impella-sheath to avoid cannula-related oozing**

- ✓ Echo-guided cannulation
- ✓ Pull-up stitching technique pushing the cannula well into the skin
- ✓ Gauze to maintain the angle of insertion

Ultrasound-guided cannulation techniques, proper positioning, and stitching with pull-up technique of the sheath to the skin are important measures to avoid local oozing or bleeding.

with tranexamic acid-soaked gauze, intranasal balloon compression, or local cauterization interventions); such bleeds can be prevented by the use of orogastric instead of nasogastric tubes. To avoid increasing the risk for thrombotic complications, reducing the anticoagulant target should be considered only when previous measures are insufficient. When indicated, the administration of peripheral UFH should be reduced. Recently, the U.S. Food and Drug Administration has approved the switch to a bicarbonate-based purge solution (25 mEq/L) in case of a contraindication to UFH or bleeding. Pump speed should be maximized when ceasing to use UFH, to minimize the risk for pump thrombosis and/or systemic embolism. Generally, the administration of reversal agents such as protamine or fresh-frozen plasma without device removal is not recommended. This reversal step should be reserved only for life-threatening bleeds. Finally, more data are needed to assess whether the increased bleeding risk outweighs the ischemic risk in patients receiving pVAD support, thus justifying less intensive antithrombotic therapy in this cohort.⁴¹ Additional measures based on the maintenance of hemostasis and monitoring of coagulation test with recommendations concerning transfusions are provided in Figure 5.⁴²

CONCLUSIONS

Bleeding and thrombotic complications can jeopardize outcomes in patients on pVAD support. Correct device and patient selection, together with a standardized approach to the prevention and treatment of microaxial pVAD-induced severe complications, are

mandatory. A comprehensive understanding of the literature concerning the mechanisms of shear-induced hemolysis and bleeding in microaxial flow pump-supported patients, as well as the different diagnostic tools, and their strengths and limitations, is vital. Combining prevention, etiologic investigation and ICU management may help mitigate complications related to bleeding and hemolysis. Prospective studies are needed to provide additional guidance to optimize and standardize hematological management in critically ill MCS-supported patients.

FUNDING SUPPORT AND AUTHOR DISCLOSURES

Drs Vandenbriele, Baldetti, Chieffo, Dauwe, Meyns, Panoulas, Papalardo, Rosenberg, Scandroglio, Schaubroeck, Schrage, Stone, Vlaselaers, and Westermann have received research and/or travel funding and speaker fees from Abiomed outside this work. Drs Vandenbriele and Dauwe are supported by a grant from University Hospitals Leuven (Klinische Onderzoeks-en Opleidingsraad). Dr Beneduce has received speaker fees from Boston Scientific. Dr Chieffo has received consultant and speaker fees from Biosensor, Boston Scientific, Medtronic, Menarini, and Shockwave Medical. Dr Gorog has received speaker honoraria from AstraZeneca; and has received institutional research grants from Bayer, Medtronic, AstraZeneca, Werfen and Alpha MD. Dr Huber has received honoraria for consulting and lecturing from Amgen, AstraZeneca, Bayer, Boehringer Ingelheim, Bristol Myers Squibb, Chiesi, Daiichi Sankyo, Novartis, Pfizer, and Sanofi. Dr Pöss has received research funding from the German Cardiac Society, the German Foundation of Heart Research, and the Dr Rolf M. Schwiete Foundation. Dr Schrage has received speaker fees from AstraZeneca. Dr Stone has received speaker honoraria from Medtronic, Pulnovo, Infraredx, and Abbott; has served as a consultant to Daiichi Sankyo, Valfix, TherOx, Robocath, HeartFlow, Ablative Solutions, Vectorious, Miracor, Neovasc, Ancora, Elucid Bio, Occlutech, CorFlow, Apollo Therapeutics, Impulse Dynamics, Cardiomech, Gore, Amgen, Adona Medical, Millennia Biopharma; and has equity or options in Ancora, Cagent, Applied Therapeutics, the Biostar family of funds, SpectraWave, Orchestra Biomed, Aria, Cardiac Success, Valfix, and Xenter. Dr Stone's daughter is an employee of IQVIA. Dr Stone's

employer, Mount Sinai Hospital, has received research support from Abbott, Abiomed, Bioventrix, Cardiovascular Systems, Philips, Biosense Webster, Shockwave, Vascular Dynamics, Pulnovo, and V-Wave. Dr Vranckx has received personal fees from Bayer, Daiichi Sankyo, CLS Behring, Pfizer-Bristol Myers Squibb, and Novartis. All other authors have reported that they have no relationships relevant to the contents of this paper to disclose.

ADDRESS FOR CORRESPONDENCE: Dr Christophe Vandenbriele, University Hospitals Leuven, Department of Cardiovascular Diseases, Herestraat 49, 3000 Leuven, Belgium. E-mail: christophe.vandenbriele@uzleuven.be.

REFERENCES

1. Amin AP, Spertus JA, Curtis JP, et al. The evolving landscape of Impella use in the United States among patients undergoing percutaneous coronary intervention with mechanical circulatory support. *Circulation*. 2020;141(4):273-284.
2. Chieffo A, Dudek D, Hassager C, et al. Joint EAPCI/ACVC expert consensus document on percutaneous ventricular assist devices. *Eur Heart J Acute Cardiovasc Care*. 2021;10(5):570-583.
3. Balthazar T, Vandenbriele C, Verbrugge FH, et al. Managing patients with short-term mechanical circulatory support: JACC review topic of the week. *J Am Coll Cardiol*. 2021;77(9):1243-1256.
4. Dhruva SS, Ross JS, Mortazavi BJ, et al. Association of use of an intravascular microaxial left ventricular assist device vs intra-aortic balloon pump with in-hospital mortality and major bleeding among patients with acute myocardial infarction complicated by cardiogenic shock. *JAMA*. 2020;323(8):734-745.
5. Miller PE, Bromfield SG, Ma Q, et al. Clinical outcomes and cost associated with an intravascular microaxial left ventricular assist device vs intra-aortic balloon pump in patients presenting with acute myocardial infarction complicated by cardiogenic shock. *JAMA Intern Med*. 2022;182(9):926-933.
6. Vandenbriele C, Tavazzi G, Price S. Caution with conclusions and context of mechanical circulatory devices. *JAMA Intern Med*. 2023;183(3):277.
7. Badiye AP, Hernandez GA, Novoa I, Chaparro SV. Incidence of hemolysis in patients with cardiogenic shock treated with Impella percutaneous left ventricular assist device. *ASAIO J*. 2016;62(1):11-14.
8. Chieffo A, Ancona MB, Burzotta F, et al. Observational Multicentre Registry of Patients Treated With Impella Mechanical Circulatory Support Device in Italy: the IMP-IT registry. *Euro Intervention*. 2020;15(15):e1343-e1350.
9. Gorog DA, Price S, Sibbing D, et al. Antithrombotic therapy in patients with acute coronary syndrome complicated by cardiogenic shock or out-of-hospital cardiac arrest: a joint position paper from the European Society of Cardiology (ESC) Working Group on Thrombosis, in association with the Acute Cardiovascular Care Association (ACCA) and European Association of Percutaneous Cardiovascular Interventions (EAPCI). *Eur Heart J Cardiovasc Pharmacother*. 2021;7(2):125-140.
10. Rother RP, Bell L, Hillmen P, Gladwin MT. The clinical sequelae of intravascular hemolysis and extracellular plasma hemoglobin: a novel mechanism of human disease. *JAMA*. 2005;293(13):1653-1662.
11. Fervenza FC, Croatt AJ, Bittar CM, et al. Induction of heme oxygenase-1 and ferritin in the kidney in warm antibody hemolytic anemia. *Am J Kidney Dis*. 2008;52(5):972-977.
12. Wagener FA, Eggert A, Boerman OC, et al. Heme is a potent inducer of inflammation in mice and is counteracted by heme oxygenase. *Blood*. 2001;98(6):1802-1811.
13. Meegan JE, Shaver CM, Putz ND, et al. Cell-free hemoglobin increases inflammation, lung apoptosis, and microvascular permeability in murine polymicrobial sepsis. *PLoS ONE*. 2020;15(2):e0228727.
14. Buerck JP, Burke DK, Schmidtke DW, Snyder TA, Papavassiliou DV, O'Rear EA. Production of erythrocyte microparticles in a sub-hemolytic environment. *J Artif Organs*. 2021;24(2):135-145.
15. Roberts N, Chandrasekaran U, Das S, Qi Z, Corbett S. Hemolysis associated with Impella heart pump positioning: In vitro hemolysis testing and computational fluid dynamics modeling. *Int J Artif Organs*. 2020;43(11):710-718.
16. Jurmann MJ, Siniawski H, Erb M, Drews T, Hetzer R. Initial experience with miniature axial flow ventricular assist devices for postcardiotomy heart failure. *Ann Thorac Surg*. 2004;77(5):1642-1647.
17. Esposito ML, Morine KJ, Annamalai SK, et al. Increased plasma-free hemoglobin levels identify hemolysis in patients with cardiogenic shock and a trans valvular micro-axial flow pump. *Artif Organs*. 2019;43(2):125-131.
18. Chalmers C, Russell WJ. When does blood haemolyse? A temperature study. *Br J Anaesth*. 1974;46:742-746.
19. Nakamura M, Imamura T, Fukui T, et al. Impact of the angle between aortic and mitral annulus on the occurrence of hemolysis during Impella support. *J Artif Organs*. 2020;23(3):207-213.
20. Nakamura M, Imamura T, Hida Y, Kinugawa K. Pulmonary artery pulsatility index and hemolysis during Impella-incorporated mechanical circulatory support. *J Clin Med*. 2022;11(5):1206.
21. Vandenbriele C, Vanassche T, Price S. Why we need safer anticoagulant strategies for patients on short-term percutaneous mechanical circulatory support. *Intensive Care Med*. 2020;46(4):771-774.
22. Vandenbriele C, Arachchilage DJ, Frederiks P, et al. Anticoagulation for percutaneous ventricular assist device-supported cardiogenic shock: JACC review topic of the week. *J Am Coll Cardiol*. 2022;79(19):1949-1962.
23. Baldetti L, Beneduce A, et al. Impella malrotation within the left ventricle is associated with adverse in-hospital outcomes in cardiogenic shock. *J Am Coll Cardiol Interv*. 2023;16(6):739-741.
24. Kanji R, Vandenbriele C, Arachchilage DRJ, Price S, Gorog DA. Optimal tests to minimise bleeding and ischaemic complications in patients on extracorporeal membrane oxygenation. *Thromb Haemost*. 2022;122(4):480-491.
25. Lüsebrink E, Kellnar A, Krieg K, et al. Percutaneous transvalvular microaxial flow pump support in cardiology. *Circulation*. 2022;145:1254-1284.
26. Dufour N, Radjou A, Thuong M. Hemolysis and plasma free hemoglobin during extracorporeal membrane oxygenation support: from clinical implications to laboratory details. *ASAIO J*. 2020;66(3):239-246.
27. Extracorporeal Life Support Organization. ELSO Guidelines for Cardiopulmonary Extracorporeal Life Support. Version 1.3, November 2013. Accessed January 24, 2023. <http://www.else.org/portals/0/igd/archive/filemanager/929122ae88cusersshydocumentselsoguidelinesgeneralalleclsversion1.3.pdf>
28. O'Neill WW, Schreiber T, Wohns DHW, et al. The current use of Impella 2.5 in acute myocardial infarction complicated by cardiogenic shock: results from the USpella Registry. *J Interv Cardiol*. 2014;27(1):1-11.
29. Kormos RL, Antonides CFJ, Goldstein DJ, et al. Updated definitions of adverse events for trials and registries of mechanical circulatory support: a consensus statement of the mechanical circulatory support academic research consortium. *J Heart Lung Transplant*. 2020;39(8):735-750.
30. Thygesen K, Alpert JS, White HD. Universal definition of myocardial infarction. *Circulation*. 2007;116(22):2634-2653.
31. Cutlip DE, Windecker S, Mehran R, et al. Clinical end points in coronary stent trials: a case for standardized definitions. *Circulation*. 2007;115(17):2344-2351.
32. Mega JL, Braunwald E, Mohanavelu S, et al. Rivaroxaban versus placebo in patients with acute coronary syndromes (ATLAS ACS-TIMI 46): a randomised, double-blind, phase II trial. *Lancet*. 2009;374(9683):29-38.
33. Schulman S, Kearon C. Definition of major bleeding in clinical investigations of anti-hemostatic medicinal products in non-surgical patients. *J Thromb Haemost*. 2005;3(4):692-694.

34. Mehran R, Rao SV, Bhatt DL, et al. Standardized bleeding definitions for cardiovascular clinical trials: a consensus report from the bleeding academic research consortium. *Circulation*. 2011;123(23):2736–2747.
35. Généreux P, Piazza N, Alu MC, et al. Valve Academic Research Consortium 3: updated endpoint definitions for aortic valve clinical research. *J Am Coll Cardiol*. 2021;77(21):2717–2746.
36. Bergmark BA, Kamphuisen PW, Wiviott SD, et al. Comparison of events across bleeding scales in the ENGAGE AF-TIMI 48 trial. *Circulation*. 2019;140(22):1792–1801.
37. Salas de Armas I, Bergeron A, Bhardwaj A, et al. Surgically implanted Impella device for patients on Impella CP support experiencing refractory hemolysis. *ASAIO J*. 2022;68(12):e251–e255.
38. Berra L, Pincioli R, Stowell CP, et al. Autologous transfusion of stored red blood cells increases pulmonary artery pressure. *Am J Respir Crit Care Med*. 2014;190(7):800–807.
39. Satoh T, Xu Q, Wang L, Gladwin MT. Hemolysis-mediated toxicity during cardiopulmonary bypass ameliorated by inhaled nitric oxide gas. *Am J Respir Crit Care Med*. 2018;198(10):1244–1246.
40. Hei FL, Irou SY, Ma J, Long C. Plasma exchange during cardiopulmonary bypass in patients with severe hemolysis in cardiac surgery. *ASAIO J*. 2009;55(1):78–82.
41. Baldetti L, Nardelli P, Ajello S, et al. Anti-thrombotic therapy with cangrelor and bivalirudin in venoarterial extracorporeal membrane oxygenation patients undergoing percutaneous coronary intervention: a single-center experience. *ASAIO J*. 2023;69(7):e346–e350. <https://doi.org/10.1097/MAT.0000000000001871>
42. Helms J, Iba T, Connors JM, et al. How to manage coagulopathies in critically ill patients. *Intensive Care Med*. 2023;49:273–290.

KEY WORDS bleeding, hemolysis, management, pVAD

 **APPENDIX** For an interactive version of the Central Illustration, please see the online version of this paper.